

HEAD / NECK CTA PHANTOM

Age
Category

Adult

Body
Region

Head

Target
Modality

CT

Diagnostic
Features

Vasculature


This phantom simulates a contrast medium enhanced head in arterial phase (CT angiography).

The neck and upper thorax are included up to the aortic arch and the fifth thoracic vertebra. The vertex is included until approximately 0.5 cm below skin level. The thoracic phantom section excludes shoulder and back parts that are typically not part of CTA examinations.


The internal carotid artery has calcifications on both sides with moderate stenosis on the right side. Otherwise, the phantom has no significant vascular pathologies.

The phantom can be used in CT (including CBCT) to evaluate and optimize CTA imaging performance and post-processing applications such as vessel segmentation, including AI-enabled applications. It is also nicely suited for training purposes.

The phantom provides a detailed and realistic simulation of vascular structures, soft and bone tissues. Air voids are filled with a cellulose-polymer composite of approx. -160 HU.



HEAD / NECK CTA PHANTOM




Specifications

Size	Approx. 185 x 215 x 322 mm
Weight	Approx. 5630 g
Base material	Cellulose-polymer composite
Optimal tube voltage	120 kVp (cf page 3) - adaptable upon request -

Diagnostic features

Realistic simulation of head and neck vessels up to the aortic arch, bone and soft tissues.

Calcifications of the internal carotid artery at the carotid bifurcation on both sides with mild stenosis on the right side.



For more information visit
www.phantomx.de

HEAD / NECK CTA PHANTOM


General indications

- The phantom is made of a cellulose-polymer composite material with properties similar to hardwood. If handled carefully, it will last a long time.
- The phantom is coated with a protective layer. If the protective layer is undamaged, the phantom can be cleaned using a damp cloth (water or mild detergent).
- Protect from direct sunlight.
- Maintain a storage temperature of 10 °C to 30 °C. If the phantom is exposed to temperatures below -10 °C or above 45 °C, it can be severely damaged.
- The phantom is not equipped for dose measurements with dosimeters and it is not suited for material characterization with dual energy CT.
- The phantom is not certified as medical device.
- Air voids are filled with cellulose-polymer composite of approx. -160 HU.


Attenuation properties

Soft Tissue

Linear attenuation coefficients [cm⁻¹] (calculated)




Hounsfield units (calculated)




Bone Tissue

Linear attenuation coefficients [cm⁻¹] (calculated)



Hounsfield units (calculated)



Tissue Reference: Woodard HQ, White DR. The composition of body tissues. Br J Radiol. 1986.